

S-Adenosyl-L-Homocysteine (SAH) ELISA Kit 2 (For Plasma, Serum or Tissue Samples)

Catalog No. IK00302s

Packaging size 96 tests

Detection range 31.25 nM – 2000 nM

Target details

Methionine is an amino acid with sulfur methyl group, which is converted into S-adenosylmethionine (SAM) with ATP by methionine adenosyltransferase (EC 2.5.1.6). SAM is the sole methyl group donor for over 50 kinds of biologically active substances such as DNA, RNA, protein, phospholipids, hormones and neuro-transmitters, etc. After methyl group is transferred away from SAM under a methyltransferase, S-adenosylhomocysteine is formed, which is further metabolized into homocysteine (Hcy) after adenosine is removed. The homocysteine is metabolized to methionine through N5-methyl tetrahydrofolate methyltransferase (EC1.1.1.68) and coenzyme vitamin B12 catalytic pathway accepting methyl from N5-methyltetrahydrofolate. This is methionine cycle.

Methylation index is defined as the ratio of SAM and SAH. Methylation index is a better marker for methylation status and methylation capability.

SAH and Hcy connect the processes of providing methyl group to critical biological molecules and recycling methyl group from N-5-methyltetrahydrofolate. The levels of SAH will determine or reflect whether the methionine cycle is normal or not. Therefore, finding a way to better quantify them has valuable practical implications. Elevated SAH can cause damage to vascular endothelial cells, which is related to the levels of global genome methylation. SAH may be a potential biomarker for atherosclerosis.

This immunoassay kit is used to measure the level of SAH in samples of soluble form.

Principle of the assay

This direct competitive ELISA (Enzyme Linked Immunosorbent assay) is designed to measure the level of S-adenosyl-L-methionine (SAH) in the sample. SAH conjugated with macromolecule is immobilized on the micro-titer plate. Standards and samples are pipetted into the wells, and then the HRP-conjugated antibody against SAH is added. The free SAH molecule in samples or standards competes with the immobilized SAH on the micro-titer plate surface for binding sites of the antibody. After discarding the mixed solution and washing each well, TMB substrate solution is added. The substrate solution turns blue under the effect of HRP (horseradish peroxidase), and changes into yellow once stop solution (acid) is added. The color develops in inverse proportion to the amount of SAH in the sample (or standards). The optical density of the remaining solution (OD450) is measured at 450nm using micro-plate spectrophotometer. The level of SAH in samples can be calculated through standard curve generated with standards.

Sample collection and storage

1. Sample collection must be carried out at 4°C. Remove precipitation by centrifugation when necessary, and test sample as soon as possible or store at -20°C or below. Avoid repeated freeze-thaw cycles.
2. Avoid NaN₃ in samples, since it will inactivate Horseradish peroxidase (HRP).
3. The kit is used for plasma, serum or tissue/cell homogenate samples.

Procedure

1. Rewarm all reagents to room temperature and mix them well. Take out appropriate number of wells and put the remaining wells back into the Ziploc. Seal the Ziploc and stored it at -20°C.
2. Add 50ul sample diluent (blank well), standards and sample into each well.
3. Prepare HRP-conjugated antibody solution: Dilute HRP-antibody with HRP-antibody diluent at 1:500, which should be used up within a week. If it will be used after a week, please prepare it when needed. Mix thoroughly and store in dark.
4. Add 50ul HRP-conjugated antibody into each well except for the blank well.
5. Put the plate onto oscillator and shake for a while to mix the reagents, and seal the plate with micro-plate sealer. Incubate the plate at 37°C for 1 hour.
6. Peel the sealer carefully, and discard the remaining solution in the wells. Add at least 300ul wash solution in each well and maintain this state for 30 seconds, then remove wash solution. Repeat these steps 3 times to finish washing process. Or use auto-washer instead.
7. Add TMB Substrate and add 100ul blending substrate to each well. Shake gently and seal the plate. Incubate the plate at 37°C for 15 minutes without light.
8. Add 50ul Stop Solution to end reaction.
9. Measure absorbance of each well at 450 nm wavelength within 15 minutes. Set zero according to the blank well.

Components

Wash Solution (20×)	1×20ml
Assay Plate	12 ×8 wells (96-well)
HRP-antibody	1×15ul
HRP-antibody Diluent	1×7.5ml
Sample Buffer	1×0.5ml
TMB Substrate	1×11ml
Stop Solution	1×6ml
Microplate Sealers	2
S0 SAH Standard 0 nM	1×250ul
S1 SAH Standard 31.25 nM	1×250ul
S2 SAH Standard 125 nM	1×250ul
S3 SAH Standard 250 nM	1×250ul
S4 SAH Standard 500 nM	1×250ul
S5 SAH Standard 1000 nM	1×250ul
S6 SAH Standard 2000 nM	1×250ul
QC SAH Quality Control 320-480nM	1×250ul

Standard curve




Figure 1 Standard curve in competitive ELISA with SAH standard from this kit $LOGIT = \ln(A/A_{S0}/(1-A/A_{S0}))$

Data processing

Always clear the background by subtracting the absorbance at 450 nm (OD450) of blank well from that of the test well.

The binding rate of each well (standard or sample) is equal to A/A_{S0} , with A being the average absorbance of the standard wells or the sample wells, A_{S0} being the average absorbance of the S0 standard wells.

Creation of standard curve: Construct a standard curve by plotting the binding rate of each standard on the y-axis against the logarithm of its concentration on x-axis. Use quadratic polynomial curve to fit the data ($r > 0.99$). The SAH level of the sample can be calculated by substituting its binding rate into the standard curve equation. Then multiply the extent of dilution if the sample has been diluted.

Note

1. Using reagents and samples without rewarming or ambient temperature is less than 20°C may lead to reduced OD450 values.
2. Extra drying post washing may have negative effect on the results, such as poor standard curve and repeatability. To avoid this, operate the next step immediately after washing.
3. Mix solution well and wash completely, as these procedures will have influence on the assay.
4. Seal the plate with micro-plate sealer and avoid light when incubating the plate.
5. Duplicate wells for standards and samples are recommended, and quadratic polynomial is suggested for standard curve fitting ($r > 0.99$). The detected concentration of quality control vial should be in the detection range.
6. The concentrated wash solution may crystallize. Warm it up to allow salts dissolved completely before diluting.
7. The micro-plate sealers should be disposable in order to avoid cross contamination.
8. The Substrate should be kept out of light.
9. The Stop Solution is diluted sulfuric acid. Avoid direct contact with skin and other things.

Storage and valid date

Storage condition: Except for the HRP substrate and stop solution that are stored at 2-8°C, all other ingredients and strips can be frozen stored.

Expiration: About 6 months from the date of shipment when storing it in refrigerator. If not planning to use it soon, users can put the assay plate, HRP-anti-SAM antibody, standards, HRP antibody and sample diluents in freezer for longer storage time or/and better results.